

CTA CHEST/Abd/Pelvis GO UP

Indications	trauma, acute aortic syndrome, suspected aneurysm/dissection					
Diagnostic Task	Detect aneurysms, aortic dissections					
Scan mode	Helical					
Position/Landmark	Head first-Supine 1cm to shoulders/inspiration					
Topogram	AP 15mA 110kV					
kVp/Reference mass	110kv 92mas					
Rotation time/pitch	without 0.8/1.2					
Detector Configuration	32x0.7					
Table Speed/Increment	26.88					
Dose reduction	CareDose 4D					
Allowed CTDI ranges*	7mGy-50mGy					
XR29 Dose Notification value	50mGy					
Helical Set 1 NON CONTRAST	recon	body part	thickness spacing	kernel	window	recon destination
	1	Chest	1.5mmx1.5mm	Bv40	mediastinum	pacs
	if patient under 40 ask about non contrast images					
Helical Set 2 ARTERIAL	recon	body part	thickness spacing	kernel	window	recon destination
	1	CAP thin	0.6mmx0.6mm	Bv40	mediastinum	pacs/TR
	2	CAP	2mmx2mm	Bv40	mediastinum	pacs/TR
	3	Cor Chest	2mmx2mm	Bv40	mediastinum	pacs
	4	Sag Chest	2mmx2mm	Bv40	mediastinum	pacs
	5	Cor Abd	2mmx2mm	Bv40	mediastinum	pacs
	6	Sag Abd	2mmx2mm	Bv40	mediastinum	pacs
	7	Mip Lung	10mmx1mm	Bv36	Lung	pacs
	8	MIP sag aorta	5mmx2mm	Bv40	mediastinum	pacs
	9	MIP cor aorta	5mmx2mm	Bv40	mediastinum	pacs
Helical Set 1 Delay 60sec	recon	body part	thickness spacing	kernel	window	recon destination
	1	Chest	1.5mmx1.5mm	Bv40	mediastinum	pacs
Scan start/End location	NC and delay see high level summary// arterial 2cm superior to lung apices					
DFOV	NC and delay see high level summary// lesser trochanters 40cm decrease appropriately					
IV contrast volume/type	<200ml 100ml isovue 370 >200ml 125ml isovue 370 3-4cc/sec					
Scan delay	Bolus Tracking at descending aorta(level just inferior to carina) Trigger is +100HU					
<p>Comments: Being able to locate the descending aorta is important. The monitoring phase will not trigger properly and the scan will not start correctly if the roi is not placed on the correct anatomy</p>						

Patient size	weight(kg)	weight(lbs)	CTDIvol(mGy)
SMALL	50-70	110-155	4-10
AVERAGE	70-90	155-200	8-16
LARGE	90-120	200-265	14-22

NOTE*

*The AAPM recommended NEMA XR29 Dose Notification Value for an adult torso is 50mGy. Dose Notification levels less than the AAPM recommended can be set. The maximum CTDI vol should match the dose notification value. Exams with CTDI vol values less than the minimum allowed range should not be performed unless approved by a radiologist.

