



Deep Brain Stimulator WO

Reviewed By: Justin Waters, MD

Last Reviewed: August 2024

Contact: (866) 761-4200, Option 1

Sequences:

- Ax T2 1mm isotropic (TR 6000, TE 90)
- Ax T2 3D FLASH 1mm
- Ax MPR
- Coronal recon MPR
- Axial recon MPR